

Unacceptable Error: Clear Proof

Data File Path: D:\MSDCOR\LULR\04071
 Data File Name: 1707474.D
 Operator: JH
 Date Acquired: 7/24/2008 13:28
 Acq. Method File: MAN07
 Sample Name: 170707 00474 1x
 Vial Number: 4
 Calibration File: Quantification TIE (2 points)
 Last Calibration Update: Mon Jul 28 11:18:48 2008

#	Peak Label	Ret. Time	Height	Name	Target Response	Amount	Status
17	17T10	20.48	307.3	Epitestosterone	307234	5.20	Agree
21	17T10	18.30	432.4	testosterone	3513238	61.37	Agree

Control du rapport TIE
 Surface: Concentration
 15.4 11.8

Name	Target Response	Amount
17méthyltestostérone	8595490	100.00
Epitestosterone	307234	5.20
testosterone	3513238	61.37

Page 1 of 1
 D:\MSDCOR\LULR\04071\MAN07.CRT
 7/24/2008 3:17 PM
 USADA 0092

Data File Path: D:\MSDCOR\LULR\02171
 Data File Name: 1707474.D
 Operator: JH
 Date Acquired: 7/22/2008 18:02
 Acq. Method File: MAN07
 Sample Name: 170707 00474 1x
 Vial Number: 10
 Calibration File: Quantification du rapport TIE
 Last Calibration Update: Mon Jul 24 12:14:34 2008

#	Peak Label	Ret. Time	Height	Name	Target Response	Amount	Status
17	17T10	21.10	307.3	Epitestosterone	244818	17.59	Agree
21	17T10	18.30	432.4	testosterone	2621497	172.23	Agree

Control du rapport TIE
 Surface: Concentration
 10.7 5.8

Name	Target Response	Amount
Méthyltestosterone	1397200	100.00
Epitestosterone	244818	17.59
testosterone	2621497	172.23

Page 1 of 1
 D:\MSDCOR\LULR\02171\MAN07.CRT
 7/24/2008 1:07 PM
 USADA 0212

Here are the relevant screen shots.

By the laboratory's own standards, its testing was unacceptable.